Instructions for use

MT02 L REF 055001

Distributed by:
Faro S.p.A.
Via Faro, 15
20876 Ornago (MB)
Italy
Phone +39 039.6878.1
Email: info@faro.it
www.faro.it

Legal manufacturer:
KaVo Dental GmbH
Bismarckring 39
88400 Biberach
Germany
www.kavo.com

CE 0124

Table of contents

1 User instructions 4

2 Safety 7
2.1 Infection hazard 7
2.2 Improper use 7
2.3 Technical condition 7
2.4 Accessories and combination with other equipment 8
2.5 Qualification of personnel 8
2.6 Service and repair 8

3 Description of the product 10
3.1 Intended use 10
3.2 Technical Specifications 10
3.3 Transportation and storage conditions 11

4 Startup and shut down 12
4.1 Installing the MULTIflex coupling 12
4.2 Checking the amount of water 13
4.3 Checking the pressures 13
4.4 Checking the O-rings 14

5 Operation 15
5.1 Attaching the medical device 15
5.2 Removing the medical device 15
5.3 Inserting the dental bur 16
5.4 Removing the dental bur 17

6 Troubleshooting 18
6.1 Changing O-rings on the coupling 18
6.2 Cleaning the spray nozzle 18

7 Reprocessing steps in accordance with ISO 17664 20
7.1 Preparations at the site of use 20
7.2 Manual reprocessing 20
7.2.1 Manual external cleaning 20
7.2.2 Manual internal cleaning 20
7.2.3 Manual external disinfection 21
7.2.4 Manual internal disinfection 21
7.2.5 Manual drying 22
7.3 Automated reprocessing 22
7.3.1 Automated internal and external cleaning and internal and external disinfection 23
7.3.2 Automated drying 23
7.4 Care products and systems - Servicing 23
7.4.1 Servicing with KaVo Spray 24
7.4.2 Servicing with KaVo QUATTROcare PLUS 24
7.5 Packaging 25
7.6 Sterilisation 25
7.7 Storage 26

8 Optional aids and consumables 27

9 Terms and conditions of warranty 28

1 User instructions

Dear user,
Congratulations on purchasing this quality product. By following the notes below you will be able to work smoothly, economically and safely.

© Copyright by KaVo Dental GmbH

All mentioned trademarks are property of their respective owners.

KaVo Original Factory Repair

In the event of a repair, please ship your product to the KaVo Original Factory Repair using www.kavobox.com/faro.

FARO Technical Service
+39 039 6878245
service@faro.it

KaVo Technical Service
If you have any technical questions or complaints, please contact the KaVo Technical Service:
+49 (0) 7351 56-1000
service.instrumente@kavo.com

Target group

The instructions for use are intended for medical professionals, in particular dentists and dental practice personnel.

The section on startup is also intended for the service staff.

General marks and symbols

	See Chapter on User Instructions/Hazard Levels
	Important information for users and service technicians
▶	Action request
CE	CE- mark (European Community). A product bearing this mark meets the requirements of the applicable EC directives.
MD	Medical device, labelling of medical devices
135 °C	Sterilisable by steam
	Thermodisinfectable

Information on the packaging

REF	Material number
SN	Serial number
	Manufacturer
	Note: Please note accompanying documents
	Follow the electronic instructions for use
	HIBC Code
	CE mark for medical devices
	EAC conformity mark (Eurasian Conformity)
MD	Medical device, labelling of medical devices
	Transportation and storage conditions (temperature range)
	Transportation and storage conditions (air pressure)
	Transportation and storage conditions (Humidity)
	Protect from moisture (Keep dry)
	Protect from impact
	Original language German

Hazard levels

The warning and safety notes in this document must be observed to prevent personal injury and material damage. The warning notes are designated as shown below:

⚠ HAZARD

In cases which – if not prevented – directly lead to death or severe injury.

⚠ WARNING

In cases which – if not prevented – can lead to death or severe injury.

⚠ CAUTION

In cases which – if not prevented – can lead to minor or moderate injury.

CAUTION

In cases which – if not prevented – can lead to material damage.

2 Safety

NOTE
All serious events occurring in relation to the product must be reported to the manufacturer and the competent authority of the member state, in which the user and/or patient resides.

The instructions for use are an integral part of the product and must be read carefully prior to use and must be accessible at all times.
The device may only be used in accordance with the intended use, any other type of use is not permitted.

Individual warning notes must be observed in the corresponding chapters.

2.1 Infection hazard

Patients, users or third parties could be infected by contaminated medical devices.

- Take suitable personal protective measures.
- Follow the instructions for use of the components.
- Before initial startup and after each use, reprocess the product and accessories appropriately.
- Carry out the reprocessing as described in the instructions for use. The procedure has been validated by the manufacturer.
- If you deviate from this validated procedure, make sure that the reprocessing procedure is effective.
- Reprocess the product and accessories appropriately before disposal.
- In the case of injury to soft tissue, do not continue treatment in the oral cavity with instruments driven by compressed air.
- Use gloves or a finger guard when you test, insert and remove the dental bur.

2.2 Improper use

The improper use of the device could lead to burns or injuries.

- Check the technical condition before each use.

Also refer to:
2.3 Technical condition, Page 7

- Never press the push-button during operation of the device.
- Never use the instrument to keep the cheek, tongue or lip at a distance.
- Never touch soft tissue with the handpiece head or instrument cover.
- Do not use the medical device as a light probe.
- Use an appropriate light probe for illumination of the oral cavity or site of preparation.
- After treatment, place the medical device properly in the cradle without the tool.

During the preparation of abutments, heat transmission can cause thermal damage to the jawbone.

- During the preparation of abutments, make sure that the preparation times are short and that there is sufficient cooling.

2.3 Technical condition

If damaged, the device or components could injure patients, users and third parties.

- Only operate devices or components if they show no signs of damage on the outside.

- Check to make sure that the device is working properly and is in satisfactory condition before each use.
- Have parts with sites of breakage or surface changes checked by the service personnel.
- If the following defects occur, stop working and have the service personnel carry out repair work:
  - Malfunctions
  - Damage
  - Irregular running noise
  - Excessive vibration
  - Overheating
  - Dental bur is not seated firmly in the handpiece

To ensure optimum function and to prevent property damage, please comply with the following instructions:

- Service the medical device regularly with care products and systems as described in the instructions for use.
- The device should be reprocessed and stored in a dry location, according to instructions, if it is not to be used for an extended period of time.

2.4 Accessories and combination with other equipment

Use of non-authorised accessories or non-authorised modifications of the device could lead to injury.

- Only use accessories that have been approved for combination with the product by the manufacturer.
- Only use accessories that are equipped with standardised interfaces.
- Do not make any modifications to the device unless these have been approved by the manufacturer of the product.
- Use original KaVo spare parts only.

2.5 Qualification of personnel

Application of the product by users without the appropriate medical training could injure patients, users or third parties.

- Make sure that the user has read and fully comprehends the instructions for use.
- Make sure that the user has read and comprehends the national and regional regulations.
- The device may be used only if the user has completed the appropriate medical training.

2.6 Service and repair

Repairs, servicing and safety checks may only be performed by trained service personnel. The following persons are authorised to do this:

- Service technicians of KaVo branches after the appropriate product training

Observe all the following items during servicing work:

- Have the service and testing tasks carried out according to the Medical Device Operator Ordinance.
- After servicing, interventions on and repairs of the device and before re-use, have the service personnel perform safety checks on the device.
- Following expiry of the warranty, have the tool holding system checked once a year.
- KaVo recommends specifying in-house service intervals where the medical device is brought to a professional shop for cleaning, servicing and a function check. Define the service interval depending on the frequency of use.

As a result of the use of NON-KaVo original spare parts during the repair, parts such as covers may become undone and injure the patient, user or other people. This may result in aspiration, swallowing of parts and possibly even a risk of suffocation.

- Only use spare parts that comply with the specification for repair; original KaVo spare parts comply with the specification.

NOTE

If a repair is done with NON-KaVo original spare parts, this may constitute a product modification that leads to the loss of CE conformity. In the event of damage, the responsibility is with the service company or the operator.

The introduction of a modified product in the market, in which the reasonable suspicion exists to endanger the safety and health of patients or users, is prohibited by Medical Device Law §4, Abs.1 No. 1 and therefore requires its own conformity check.

3 Description of the product

MT02 L REF 055001

3.1 Intended use

Indications for use:

This medical device is:

- Intended only for dental treatment by a dental professional, the product must not be modified or used for any other purpose since this may be hazardous
- The medical device is intended for the following applications:
  - Removal of carious material
  - Removal of fillings
  - Processing of tooth and restoration surfaces
  - Tooth, cavity and crown preparations
- A medical device according to relevant national statutory regulations

Proper use:

According to these regulations, this product may only be used for the described application by a properly trained user. You need to comply with the following:

- the applicable health and safety regulations
- the applicable accident prevention regulations
- these Instructions for use

According to these regulations, the user is required:

- to only use equipment that is operating correctly
- to comply with the specified intended use
- to protect him or herself, the patient and third parties from hazards
- to prevent contamination from the product

3.2 Technical Specifications

Head size	Standard
Drive pressure	2.1 to 3.0 bar (30 to 44 psi)
Recommended drive pressure	2.8 bar (40 psi)
Return air pressure	< 0.5 bar (7 psi)
Spray water pressure	0.8 to 1.0 bar (12 to 15 psi)
Spray air pressure	1.0 to 2.5 bar (15 to 36 psi)
Air consumption	45 to 55 Nl/min
Idle speed	350,000 to 450,000 min^{-1}

Recommended application force	2 to 3 N
Connector	MULTIflex couplings

NOTE
The air consumption and idle speed values were determined at a drive pressure of 2.8 bar (40 psi).

3.3 Transportation and storage conditions

CAUTION

Startup after refrigerated storage.
Malfunction.
- Prior to startup, strongly refrigerated products must be allowed to warm up to a temperature of 20 °C to 25 °C (68 °F to 77 °F).

	Temperature: -29 °C to +50 °C (-20 °F to +122 °F)
	Relative humidity: 5% RH to 85% RH absence of condensation
	Air pressure: 700 hPa to 1060 hPa (10 psi to 15 psi)
	Protect from moisture (Keep dry)

4 Startup and shut down

⚠ WARNING

Hazard from contaminated products.

Infection hazard to the dentist and patient.

- Prior to initial startup and after each use, reprocess the product and accessories.

⚠ WARNING

Dispose of the product in appropriate manner.

Infection hazard.

- Reprocess the product and accessories before disposal.

Also refer to:
7 Reprocessing steps in accordance with ISO 17664, Page 20

Currently applicable packaging law

Dispose of and recycle the packaging appropriately in accordance with current packaging law, employing waste management / recycling companies. Comply with the comprehensive return system. KaVo has had its packaging licensed for this purpose. Please comply with the regional public waste-disposal system.

CAUTION

Damage from soiled and moist cooling air/compressed air.
Contaminated and moist cooling air can cause malfunctions.

- Make sure that the supply of cooling air is dry, clean and uncontaminated according to DIN EN ISO 7494-2.

4.1 Installing the MULTIflex coupling

⚠ WARNING

Detaching the medical device during treatment.

A medical device that is not properly locked can detach from the coupling during treatment.

- Before each use, check if the medical device is securely locked onto the coupling.

- Screw the MULTIflex coupling onto the turbine hose and tighten it with the wrench (**KaVo Mat. no. 0.411.1563**).



- Rotate the spray ring on the MULTIflex coupling (465 series) in order to regulate the water supply.



- Turn the spray ring clockwise to reduce the water volume.
- Turn the spray ring anticlockwise to increase the water volume.

4.2 Checking the amount of water

⚠ CAUTION

Overheating of the tooth due to insufficient amount of cooling water.

Insufficient spray water can cause the medical device to overheat and damage the pulp and tooth.

- Adjust the water flow for the spray cooling to a minimum of 50 mL/min (3.1 inch^3).
- Check the spray water channels and if necessary clean the spray nozzles with the nozzle needle (**KaVo Mat. no. 1.003.8150**).



4.3 Checking the pressures

CAUTION

Damage from soiled and moist cooling air/compressed air.

Contaminated and moist cooling air can cause malfunctions.

- Make sure that the supply of cooling air is dry, clean and uncontaminated according to DIN EN ISO 7494-2.

- Insert the pressure gauge (**KaVo Mat. no. 0.411.8731**) between the coupling and the medical device and check the following pressures:
  - Drive pressure
  - Recommended drive pressure
  - Return air pressure
  - Spray water pressure
  - Spray air pressure

Also refer to:

3.2 Technical Specifications, Page 10

4.4 Checking the O-rings

CAUTION

Missing or damaged O-rings.
Malfunction and premature failure.
- Make sure that all O-rings are on the coupling and are undamaged.

Number of available O-rings: 5

5 Operation

NOTE
At the beginning of each workday, the water-conducting systems should be rinsed for at least 2 minutes (without transmission handpieces being attached) and if there is a risk of contamination from reflux or back suction, the system may also need to be rinsed for 20 to 30 seconds after each patient.

⚠ WARNING

Rotating dental bur.

Cuts, infection and burn injury.

- Never push the press-button while the dental bur is rotating.
- Do not touch the dental bur while it is rotating.
- Never touch soft tissue with the handpiece head or instrument cover.
- Remove the dental bur from the handpiece after treatment to avoid injury and infection during storage.

⚠ CAUTION

Heat transmission during the preparation of abutments.

Thermal damage to the jawbone.

- During the preparation of abutments, make sure that the preparation times are short and that there is sufficient cooling.

5.1 Attaching the medical device

⚠ WARNING

Detaching the medical device during treatment.

A medical device that is not properly locked can detach from the coupling during treatment.

- Before each use, check if the medical device is securely locked onto the coupling.

CAUTION

Inexact coupling.
Reduces the service life of the lamp.

- Avoid inexact coupling.
- Check the secure fit of the (LUX) handpieces on the coupling by pulling on them.



- Accurately attach the medical device to the MULTIflex and push it to the rear until the coupling audibly locks in the medical device.
- Pull on the medical device to make sure that it is securely affixed to the coupling.

5.2 Removing the medical device

- Grasping the coupling, twist the medical device slightly and pull it off.

5.3 Inserting the dental bur

NOTE

Only use carbide burs or diamond grinders that comply with DIN EN ISO 1797 type 3, are made of steel or hard metal and meet the following criteria:

- Shaft diameter: 1.59 to 1.60 mm (0.0626 in to 0.0629 in)
- Overall length: max. 25 mm
- Shaft clamping length: at least 11 mm
- Blade diameter: max. 2 mm (0.0787 in)

⚠ WARNING

Use of unapproved dental burs.

Injury to the patient or damage to the medical device.

- Comply with the instructions for use and the intended use of the dental bur.
- Only use dental burs that do not deviate from the specified data.

⚠ CAUTION

Do not use damaged dental burs.

Risk of injury from swallowing a dental bur that falls out.

- Do not use damaged dental burs.
- Do not use dental burs that have been hit.
- Do not use dental burs with adherent soiling.
- Do not use visibly imbalanced dental burs.

⚠ CAUTION

Dental bur with damaged, worn or deformed shafts.

Injury hazard, dental bur may fall out during treatment.

- Never use dental burs with damaged, worn or deformed shafts.

⚠ CAUTION

Defective chuck system.

Injury hazard, dental bur may fall out during treatment.

- Pull on the dental bur to check if the chuck system works properly and if the dental bur is firmly clamped.

⚠ CAUTION

Contaminated, sharp-edged dental bur.

Infections or cuts.

- Use gloves or a finger guard when you test, insert and remove the dental bur.

CAUTION

Dental bur with damaged, worn or deformed shafts.

Material damage to the chuck system, dental bur is difficult or impossible to remove from the chuck system.

- Never use dental burs with damaged, worn or deformed shafts.

CAUTION

Dental bur shaft slips inside the chuck due to excessive speed of the dental bur or abrupt engagement of the dental bur.
Material damage to dental bur shaft and chuck system, reduction of the service life of dental bur and chuck system.

- Do not operate the dental bur at a higher speed than recommended by the manufacturer.

- Firmly press the push-button with your thumb and, simultaneously, always insert the dental bur until it hits the end stop.
- Check if the dental bur is seated securely by pulling on it.

5.4 Removing the dental bur

⚠ WARNING

Rotating dental bur.

Cuts, infection and burn injury.

- Never push the press-button while the dental bur is rotating.
- Do not touch the dental bur while it is rotating.
- Never touch soft tissue with the handpiece head or instrument cover.
- Remove the dental bur from the handpiece after treatment to avoid injury and infection during storage.

CAUTION

Damage to the chucking system.
Material damage.

- Never push the press-button while the dental bur is rotating.



- After the dental bur has stopped rotating, firmly press the push-button down with your thumb and simultaneously remove the dental bur.

6 Troubleshooting

⚠ WARNING

Repair WITHOUT using KaVo original spare parts.

Parts such as the cover can come loose and cause injury. Aspiration, swallowing of parts and danger of suffocation.

- Only use spare parts that comply with the specification for repair; original KaVo spare parts comply with the specification.

NOTE

If a repair is done with NON-KaVo original spare parts, this may constitute a product modification that leads to the loss of CE conformity. In the event of damage, the responsibility is with the service company or the operator.

The introduction of a modified product in the market, in which the reasonable suspicion exists to endanger the safety and health of patients or users, is prohibited by Medical Device Law §4, Abs.1 No. 1 and therefore requires its own conformity check.

6.1 Changing O-rings on the coupling

CAUTION

Improper care of the O-rings.
Malfunction or complete failure.

- Do not use Vaseline or other grease or oil.
- Spray a lint-free cloth with KaVo Spray and apply it to the O-rings on the coupling.

- Press the O-ring between your fingers to form a loop.
- Push the O-ring to the front, and remove it.
- Spray a lint-free cloth with KaVo Spray, apply it to the new O-rings and insert them in the recesses.

6.2 Cleaning the spray nozzle

⚠ WARNING

Hazard from contaminated products.

Infection hazard to the dentist and patient.

- Prior to initial startup and after each use, reprocess the product and accessories.

⚠ CAUTION

Overheating of the tooth due to insufficient amount of cooling water.

Insufficient spray water can cause the medical device to overheat and damage the pulp and tooth.

- Check the spray water channels and if necessary clean the spray nozzles with the nozzle needle (**KaVo Mat. no. 1.003.8150**).



7 Reprocessing steps in accordance with ISO 17664

7.1 Preparations at the site of use

⚠ WARNING

Hazard from contaminated products.

Contaminated products are associated with an infection hazard.

- Take suitable personal protective measures.

⚠ WARNING

Sharp dental bur in the medical device.

Injury hazard from sharp and/or pointed dental bur.

- Remove the dental bur.

- Reprocess the medical device right after treatment.
- The medical device must be dry when transported to reprocessing.
- To minimise the risk of infection during reprocessing, always wear protective gloves.
- Remove the dental bur from the medical device.
- Remove all residual cement, composite or blood immediately.
- Do not place in solutions or similar substances.

7.2 Manual reprocessing

⚠ WARNING

Sharp dental bur in the medical device.

Injury hazard from sharp and/or pointed dental bur.

- Remove the dental bur.

CAUTION

Never reprocess this medical device in an ultrasonic cleaner.
Malfunction and material damage.

- Reprocess in a washer disinfector or by hand only.

7.2.1 Manual external cleaning

Accessories required:

- Tap water (drinking water)
- Brush, e.g. medium-hard toothbrush



- Brush off under flowing tap water at least 10 seconds.

7.2.2 Manual internal cleaning

KaVo CLEANspray is designed for validated manual internal cleaning (removal of residual protein).

- Cover the medical device with the KaVo Cleanpac bag, and place it on the corresponding care adapter.
- Hold the can vertically.
- Press the spray button three times for 2 seconds each time.

- Remove the medical device from the spray attachment and let the cleanser act for 1 minute.

Also refer to:
Instructions for use KaVo CLEANspray

- If a manual external and internal disinfection do not follow directly, dry the medical device with KaVo DRYspray.

Also refer to:
7.2.5 Manual drying, Page 22

7.2.3 Manual external disinfection

⚠ WARNING

Incomplete disinfection.

Infection hazard.

- Only use disinfection procedures that are verified to be bactericidal, fungicidal and virucidal.
- If the disinfectants/disinfection procedures fail to meet the mandatory national requirements, perform a final sterilisation with the sterilisation parameters as described.

CAUTION

Never reprocess the medical device with chloride-containing products.
Malfunction and material damage.

- Reprocess in a washer disinfector or by hand only.

KaVo recommends the following products based on the compatibility of the materials. The microbiological efficacy must be ensured by the disinfectant manufacturer and proven by an expert opinion.

Approved disinfectants:

- CaviWipes and CaviCide made by Metrex
- Mikrozid AF made by Schülke & Mayr (Liquid or wipes)
- FD 322 made by Dürr

Consumables required:

- Cloths for wiping the medical device.

- Spray the disinfectant on a cloth, then wipe down the medical device and allow the disinfectant to act according to the instructions of the disinfectant manufacturer.
- Follow the instructions for use of the disinfectant.

7.2.4 Manual internal disinfection

⚠ WARNING

Incomplete disinfection.

Infection hazard.

- Only use disinfection procedures that are verified to be bactericidal, fungicidal and virucidal.
- If the disinfectants/disinfection procedures fail to meet the mandatory national requirements, perform a final sterilisation with the sterilisation parameters as described.

CAUTION

Never reprocess the medical device with chloride-containing products.
Malfunction and material damage.

- Reprocess in a washer disinfector or by hand only.

The efficacy of manual internal disinfection must be demonstrated by the manufacturer of the disinfection agent. With KaVo products, only use disinfection agents that have been released by KaVo with respect to the compatibility of materials (e.g. WL-cid / made by ALPRO).

- Cover the medical device with the KaVo Cleanpac bag, and place it on the corresponding care adapter.
- Hold the can vertically.
- Press the spray key for at least 3 seconds.
- Remove the medical device from the spray attachment and let the disinfectant act for 2 minutes.
- Follow the instructions for use of the disinfectant.

7.2.5 Manual drying

Use KaVo DRYspray for subsequent drying of the air, water and gear unit ducts.

- Cover the medical device with the KaVo Cleanpac bag, and place it on the corresponding care adapter.
- Hold the can vertically.
- Press the spray key for at least 3 seconds.

Also refer to:
Instructions for use KaVo DRYspray

- Immediately after drying, lubricate the KaVo medical device with care agents from the KaVo care system.

Also refer to:
7.4 Care products and systems - Servicing, Page 23

7.3 Automated reprocessing

⚠ WARNING

Incomplete disinfection.

Infection hazard.

- Only use disinfection procedures that are verified to be bactericidal, fungicidal and virucidal.
- If the disinfectants/disinfection procedures fail to meet the mandatory national requirements, perform a final sterilisation with the sterilisation parameters as described.

⚠ WARNING

Sharp dental bur in the medical device.

Injury hazard from sharp and/or pointed dental bur.

- Remove the dental bur.

CAUTION

Never reprocess the medical device with chloride-containing products.
Malfunction and material damage.

- Reprocess in a washer disinfector or by hand only.

CAUTION

Never reprocess this medical device in an ultrasonic cleaner.
Malfunction and material damage.

- Reprocess in a washer disinfector or by hand only.

7.3.1 Automated internal and external cleaning and internal and external disinfection



KaVo recommends washer disinfectors in accordance with EN ISO 15883-1, which are operated using alkaline cleaning agents.

The validations were performed using a Miele washer disinfector and the "VARIO-TD" programme and a mild alkaline cleaner made by Dr. Weigert.

In addition, KaVo recommends the use of a rinsing agent.

- For programme settings as well as cleansers and disinfectants, please refer to the Instructions for Use of the washer disinfector.

7.3.2 Automated drying

The drying procedure is usually part of the cleaning programme of the washer disinfector.

NOTE
Please comply with the instructions for use of the washer disinfector.

- In order not to affect the KaVo medical device, make sure that the product is dry on the inside and outside after completion of the cycle.
- Remove any residual liquids with KaVoDRYspray.

Also refer to:
7.2.5 Manual drying, Page 22

- Immediately after drying, lubricate the KaVo medical device with care agents from the KaVo care system.

7.4 Care products and systems - Servicing

⚠ WARNING

Sharp dental bur in the medical device.

Injury hazard from sharp and/or pointed dental bur.

- Remove the dental bur.

⚠ CAUTION

Improper service and care.

Risk of injury.

- Service regularly with suitable agents.

NOTE
KaVo guarantees the proper function of KaVo products only if the care products listed by KaVo as accessories are used, since these were tested for proper use on our products.

7.4.1 Servicing with KaVo Spray

KaVo recommends servicing the product as part of the reprocessing after each use, i.e. after each cleaning, disinfection, and before each sterilisation. Though no later than after 30 minutes of operation and, in general, after 6 patients.



- Remove the dental bur from the medical device.
- Cover the medical device with the KaVo Cleanpac bag, and place it on the corresponding care adapter.
- Press the spray key for 1 to 2 seconds.

Servicing the chuck

KaVo recommends servicing the chucking system once weekly.



- Remove the dental bur from the medical device.
- Position the tip of the spray nipple in the opening, and apply the spray.
- Press the spray key for 1 to 2 seconds.

7.4.2 Servicing with KaVo QUATTROcare PLUS

Cleaning and servicing device with expansion pressure for internal cleaning of inorganic residues and optimum care.

(no validated cleaning of the interior in accordance with German Robert Koch Institute requirements)

KaVo recommends servicing the product as part of the reprocessing after each use, i.e. after each cleaning, disinfection, and before each sterilisation. Though no later than after 30 minutes of operation and, in general, after 6 patients.

- Remove the dental bur from the medical device.
- Service the device in the QUATTROcare PLUS.

Also refer to:
Instructions for use KaVo QUATTROcare PLUS

Servicing the chuck

KaVo recommends servicing the chuck system once weekly using the chuck servicing programme integrated in the device.

NOTE
Handpieces must be taken off the service couplings before the chuck service can be started and performed.

- Close the front flap and press the chuck service button for at least three seconds until the spray canister control LED flashes three times consecutively.
  ⇨ The device is in chuck service mode.
- Remove the service coupling of the chuck from the side hatch of the QUATTROcare PLUS and attach it to coupling service point four, on the far right. A MULTIflex adaptor must be mounted there.
- Press the handpiece together with the guide bush of the chuck to be serviced against the tip of the service coupling.
- Press the button marked with the chuck service symbol.



NOTE

Close the chuck service mode.

Option 1: Place the dental handpieces in the QUATTROcare PLUS 2124 A, close the front flap and start the service procedure.

Option 2: After three minutes with no service procedure running, the device automatically switches back to normal service mode.

Also refer to:
7.4.2 Servicing with KaVo QUATTROcare PLUS, Page 24

7.5 Packaging

NOTE

The sterile goods package must be large enough to accommodate the product without stretching the packaging. The quality and use of the packaging of the items to be sterilised must meet the applicable standards and be appropriate for the sterilisation process!

- Seal the medical device separately in a sterile pack.

7.6 Sterilisation

Sterilisation in a steam steriliser (autoclave) in accordance with EN 13060 / EN ISO 17665-1

⚠ CAUTION

Improper service and care.

Risk of injury.

- Service regularly with suitable agents.

CAUTION

Contact corrosion due to moisture.

Damage to product.

- Immediately remove the product from the steam steriliser after the sterilisation cycle.



The medical device has a max. temperature resistance of up to 138 °C (280.4 °F).

Sterilisation parameters:

Select a suitable process from the following sterilisation processes (depending on the available steriliser):

- Steriliser with triple pre-vacuum:
  - at least 3 minutes at 134 °C -1 °C / +4 °C (273 °F -1.6 °F / +7.4 °F)
- Steriliser using the gravity method:
  - at least 10 minutes at 134 °C -1 °C / +4 °C (273 °F -1.6 °F / +7.4 °F)
  - at least 30 minutes at 121 °C -1 °C / +4 °C (250 °F -1.6 °F / +7.4 °F)
- Remove the medical device from the steam steriliser immediately after the sterilisation cycle is completed.
- Use according to the manufacturer's Instructions for Use.

7.7 Storage

Reprocessed products must be stored appropriately protected from dust in a dry, dark and cool space with a low germ level.

NOTE
Comply with the expiry date of the sterilised items.

8 Optional aids and consumables

Available from authorised KaVo dealers.

Material summary	KaVo Mat. no.
Replacement cartridge Standard	**1.008.4163**
Replacement cartridge Mini	**1.008.4164**
Wrench for cover	**0.411.3053**
Wrench Mini	**1.006.3384**
INTRA Instrument stand	**3.005.5204**
Nozzle needle	**1.003.8150**
Spray head for MULTIflex turbine handpieces for KaVo Spray	**0.411.9921**
Spray head INTRA (incl. nipple)	**0.411.9911**
Cleanpac 10 units	**0.411.9691**
CLEANspray/DRYspray Starter set 2116 P consisting of: ▪ 1 bottle of KaVo CLEANspray ▪ 1 bottle of KaVo DRYspray ▪ 1 adapter set INTRA + MULTIflex	**1.007.0573**
Adapter KaVo MULTIflex for KaVo CLEANspray/DRYspray	**1.007.1775**
KaVo CLEANspray 2110 P	**1.007.0579**
KaVo DRYspray 2117 P	**1.007.0580**
KaVo Spray 2112 A	**0.411.9640**
QUATTROcare plus Spray 2140 P	**1.005.4525**
Service coupling BORDEN 2-hole	**1.002.1217**
Chuck servicing set	**1.003.1253**

9 Terms and conditions of warranty

This KaVo medical device is subject to the following warranty conditions:
KaVo grants the end customer a warranty of proper function and guarantees zero defects in respect of material and workmanship for a period of 18 months from the date of the invoice, subject to the following conditions:
With regard to justified complaints KaVo grants warranty in the form of a free of charge repair or delivery of a replacement. Other claims of any kind whatsoever, in particular with respect to compensation, are excluded. In the event of default, gross negligence or intent, this shall only apply insofar as this does not conflict with mandatory statutory provisions.
KaVo shall not be liable for defects and consequences thereof that have arisen or may arise from natural wear, improper handling, cleaning, servicing or maintenance, non-compliance with operating, maintenance or connection instructions, calcination or corrosion, contaminated air or water supplies or chemical or electrical factors deemed abnormal or impermissible in accordance with the KaVo instructions for use or other manufacturer's instructions. The warranty granted does, in general, not extend to lamps, optical fibres made of glass and glass fibres, glassware, rubber parts, and the colourfastness of plastic parts.
Any liability is excluded if defects or the consequences thereof are due to the customer or third parties not authorized by KaVo interfering with or modifying the product.
Warranty claims can only be asserted if proof of sale in the form of a copy of the invoice or delivery note is presented with the product. The dealer, purchase date, type, and serial number must be clearly evident from this document.

1.015.3083 · bd · 20230620 · 01 · en